

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112118\
 Data File : VU028324.D
 Acq On : 21 Nov 2018 17:01
 Operator : MD/SY
 Sample : J5997-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Nov 22 00:05:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 21 22:33:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	167865	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	157223	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	71558	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	65514	5.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.80%
7) Chloroethane-d5	1.68	69	51892	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.27	63	95157	3.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.80%
20) 2-Butanone-d5	4.19	46	185651	43.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.84%
24) Chloroform-d	4.65	84	108661	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
26) 1,2-Dichloroethane-d4	5.31	65	61644	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.34	84	222580	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.33	67	80840	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.57	98	190490	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	7.85	79	28611	5.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	109.00%
46) 2-Hexanone-d5	8.31	63	130646	41.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.00%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	55195	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	62612	4.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112118\
Data File : VU028324.D
Acq On : 21 Nov 2018 17:01
Operator : MD/SY
Sample : J5997-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Nov 22 00:05:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 21 22:33:20 2018
Response via : Initial Calibration

