

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112118\
 Data File : VU028305.D
 Acq On : 21 Nov 2018 09:27
 Operator : MD/SY
 Sample : VU1121WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK45

Quant Time: Nov 22 10:20:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 21 22:33:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	169885	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	157925	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	68609	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	67850	5.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.20%
7) Chloroethane-d5	1.68	69	52840	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.80%
11) 1,1-Dichloroethene-d2	2.27	63	98649	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.60%
20) 2-Butanone-d5	4.20	46	198676	46.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.88%
24) Chloroform-d	4.65	84	112758	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	5.31	65	62685	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
32) Benzene-d6	5.34	84	222860	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.33	67	82325	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	7.56	98	194704	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
43) trans-1,3-Dichloropropene-	7.85	79	29944	5.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	113.60%
46) 2-Hexanone-d5	8.31	63	139475	44.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.22%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	57006	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	61665	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112118\
Data File : VU028305.D
Acq On : 21 Nov 2018 09:27
Operator : MD/SY
Sample : VU1121WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK45

Quant Time: Nov 22 10:20:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 21 22:33:20 2018
Response via : Initial Calibration

