

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112118\
 Data File : VU028315.D
 Acq On : 21 Nov 2018 13:27
 Operator : MD/SY
 Sample : J6026-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0FN6

Quant Time: Nov 21 23:33:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 21 22:33:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	171920	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	163870	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	73778	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	132193	10.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	204.40%#
7) Chloroethane-d5	1.68	69	104432	9.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	187.00%#
11) 1,1-Dichloroethene-d2	2.28	63	193444	7.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	150.20%#
20) 2-Butanone-d5	4.19	46	397401	91.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	183.58%#
24) Chloroform-d	4.65	84	224846	9.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	186.20%#
26) 1,2-Dichloroethane-d4	5.31	65	127202	9.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	190.60%#
32) Benzene-d6	5.34	84	462669	9.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	196.20%#
36) 1,2-Dichloropropane-d6	6.33	67	164209	9.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	191.80%#
41) Toluene-d8	7.57	98	401665	9.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	190.20%#
43) trans-1,3-Dichloropropene-	7.85	79	60295	11.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	220.40%#
46) 2-Hexanone-d5	8.31	63	283832	86.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	173.00%#
57) 1,1,2,2-Tetrachloroethane-	10.43	84	115315	9.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	186.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	132327	9.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	192.80%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112118\
Data File : VU028315.D
Acq On : 21 Nov 2018 13:27
Operator : MD/SY
Sample : J6026-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0FN6

Quant Time: Nov 21 23:33:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 21 22:33:20 2018
Response via : Initial Calibration

