

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U112119\
 Data File : VU035804.D
 Acq On : 21 Nov 2019 15:35
 Operator : JC/SP
 Sample : K5910-16DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAQJ9DL

Quant Time: Nov 22 08:25:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 22 01:33:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	224709	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	266735	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	84851	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	156071	5.32	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.40%
7) Chloroethane-d5	1.94	69	141577	5.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.40%
11) 1,1-Dichloroethene-d2	2.59	63	226137	4.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.80%
20) 2-Butanone-d5	4.70	46	363489	53.51	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.02%
24) Chloroform-d	5.11	84	266837	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.40%
26) 1,2-Dichloroethane-d4	5.75	65	154969	5.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.00%
32) Benzene-d6	5.77	84	529636	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.00%
36) 1,2-Dichloropropane-d6	6.74	67	184563	5.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.60%
41) Toluene-d8	7.93	98	436977	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
43) trans-1,3-Dichloropropene-	8.22	79	61262	4.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.80%
46) 2-Hexanone-d5	8.70	63	251115	46.82	ug/L	0.03
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.64%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	142982	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	12.22	152	106650	5.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.80%

Target Compounds					Ovalue
34) Trichloroethene	6.58	95	254540	9.324 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU112119\
Data File : VU035804.D
Acq On : 21 Nov 2019 15:35
Operator : JC/SP
Sample : K5910-16DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAQJ9DL

Quant Time: Nov 22 08:25:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 22 01:33:44 2019
Response via : Initial Calibration

