

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
 Data File : VU052020.D
 Acq On : 22 Nov 2022 01:52
 Operator : JC/MD
 Sample : N5737-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4348

Quant Time: Nov 22 02:12:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 22 01:03:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	431176	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	425429	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	158986	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	173403	46.409	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	92.820%	
7) Chloroethane-d5	1.906	69	148315	50.121	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	100.240%	
11) 1,1-Dichloroethene-d2	2.565	63	202516	36.565	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	73.140%	
21) 2-Butanone-d5	4.613	46	236045	105.127	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	105.130%	
24) Chloroform-d	5.060	84	300087	49.018	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.040%	
26) 1,2-Dichloroethane-d4	5.700	65	190148	51.643	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.280%	
32) Benzene-d6	5.726	84	627513	49.486	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.980%	
36) 1,2-Dichloropropane-d6	6.687	67	209535	51.435	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	102.860%	
41) Toluene-d8	7.896	98	514621	45.422	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.840%	
43) trans-1,3-Dichloroprop...	8.176	79	77825	44.876	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.760%	
47) 2-Hexanone-d5	8.629	63	158087	99.479	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	99.480%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	316387	49.091	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	98.180%	
66) 1,2-Dichlorobenzene-d4	12.192	152	177374	53.950	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	107.900%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
Data File : VU052020.D
Acq On : 22 Nov 2022 01:52
Operator : JC/MD
Sample : N5737-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4348

Quant Time: Nov 22 02:12:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 22 01:03:12 2022
Response via : Initial Calibration

