

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
 Data File : VU052028.D
 Acq On : 22 Nov 2022 06:11
 Operator : JC/MD
 Sample : N5658-10ME
 Misc : 2.01g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ESQT0ME

Quant Time: Nov 22 07:20:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 22 03:27:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	558380	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	574101	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	254077	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	324665	67.097	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 134.200%		
7) Chloroethane-d5	1.871	69	113029	29.495	ug/L	-0.03
Spiked Amount	50.000	Range 70 - 130	Recovery	= 59.000%#		
11) 1,1-Dichloroethene-d2	2.543	63	416921	58.128	ug/L	-0.02
Spiked Amount	50.000	Range 60 - 125	Recovery	= 116.260%		
21) 2-Butanone-d5	4.623	46	511825	176.022	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 176.020%#		
24) Chloroform-d	5.057	84	606058	76.445	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 152.900%#		
26) 1,2-Dichloroethane-d4	5.697	65	376835	79.031	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 158.060%#		
32) Benzene-d6	5.719	84	1304518	76.234	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 152.460%#		
36) 1,2-Dichloropropane-d6	6.684	67	423689	77.070	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 154.140%#		
41) Toluene-d8	7.896	98	1184356	77.464	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 154.920%#		
43) trans-1,3-Dichloroprop...	8.179	79	178291	76.185	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 152.360%#		
47) 2-Hexanone-d5	8.645	63	442595	206.386	ug/L	0.02
Spiked Amount	100.000	Range 45 - 130	Recovery	= 206.390%#		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	737022	84.742	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 169.480%#		
66) 1,2-Dichlorobenzene-d4	12.192	152	463485	88.213	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 176.420%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
Data File : VU052028.D
Acq On : 22 Nov 2022 06:11
Operator : JC/MD
Sample : N5658-10ME
Misc : 2.01g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ESQT0ME

Quant Time: Nov 22 07:20:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 22 03:27:10 2022
Response via : Initial Calibration

