

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
 Data File : VU052013.D
 Acq On : 21 Nov 2022 22:50
 Operator : JC/MD
 Sample : N5710-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 22 01:07:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 22 01:03:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.246	114	437284	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	427845	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	153519	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	191961	50.658	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.320%
7) Chloroethane-d5	1.906	69	160700	53.548	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.100%
11) 1,1-Dichloroethene-d2	2.565	63	224208	39.916	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.840%
21) 2-Butanone-d5	4.616	46	254050	111.566	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.570%
24) Chloroform-d	5.060	84	321474	51.779	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.560%
26) 1,2-Dichloroethane-d4	5.700	65	200413	53.671	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.340%
32) Benzene-d6	5.726	84	680364	53.351	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.700%
36) 1,2-Dichloropropane-d6	6.687	67	221237	54.001	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.000%
41) Toluene-d8	7.896	98	558947	49.056	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.120%
43) trans-1,3-Dichloroprop...	8.176	79	83333	47.781	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.560%
47) 2-Hexanone-d5	8.629	63	171698	107.434	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.430%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	333536	51.459	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.920%
66) 1,2-Dichlorobenzene-d4	12.192	152	182269	57.413	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
Data File : VU052013.D
Acq On : 21 Nov 2022 22:50
Operator : JC/MD
Sample : N5710-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 22 01:07:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 22 01:03:12 2022
Response via : Initial Calibration

