

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
 Data File : VU052039.D
 Acq On : 22 Nov 2022 13:00
 Operator : JC/MD
 Sample : N5716-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 23 03:27:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 22 07:34:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	462625	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	442647	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	167921	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	192620	48.047	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.100%
7) Chloroethane-d5	1.906	69	163227	51.411	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.820%
11) 1,1-Dichloroethene-d2	2.565	63	224184	37.726	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.460%
21) 2-Butanone-d5	4.613	46	252197	104.685	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.690%
24) Chloroform-d	5.060	84	316782	48.228	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.460%
26) 1,2-Dichloroethane-d4	5.700	65	201779	51.076	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.160%
32) Benzene-d6	5.726	84	675615	51.207	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.420%
36) 1,2-Dichloropropane-d6	6.687	67	222081	52.394	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.780%
41) Toluene-d8	7.896	98	556691	47.224	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.440%
43) trans-1,3-Dichloroprop...	8.176	79	83082	46.044	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.080%
47) 2-Hexanone-d5	8.629	63	167696	101.421	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.420%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	322554	48.101	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	188025	54.147	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112122\
Data File : VU052039.D
Acq On : 22 Nov 2022 13:00
Operator : JC/MD
Sample : N5716-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 23 03:27:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 22 07:34:37 2022
Response via : Initial Calibration

