

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112123\
 Data File : VU056545.D
 Acq On : 21 Nov 2023 12:43
 Operator : MD/SY
 Sample : 05373-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 21 23:33:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 21 23:30:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	245943	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	244931	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	108786	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	108361	5.891	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.800%
7) Chloroethane-d5	1.911	69	76075	5.394	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.800%
11) 1,1-Dichloroethene-d2	2.563	65	50265	6.191	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	123.800%
20) 2-Butanone-d5	4.621	46	209872	57.215	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.440%
24) Chloroform-d	5.055	84	197525	5.551	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.000%
26) 1,2-Dichloroethane-d4	5.695	65	97670	5.940	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.800%
32) Benzene-d6	5.721	84	386717	5.299	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%
36) 1,2-Dichloropropane-d6	6.685	67	121910	5.543	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.800%
41) Toluene-d8	7.894	98	334620	5.193	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
43) trans-1,3-Dichloroprop...	8.174	79	44050	5.222	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.400%
46) 2-Hexanone-d5	8.631	63	173324	52.288	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.580%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	85721	5.455	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	118982	5.920	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.400%
Target Compounds						
16) Methylene chloride	3.046	84	10131	0.508	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112123\
Data File : VU056545.D
Acq On : 21 Nov 2023 12:43
Operator : MD/SY
Sample : 05373-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 21 23:33:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 21 23:30:53 2023
Response via : Initial Calibration

