

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112123\
 Data File : VU056565.D
 Acq On : 21 Nov 2023 21:36
 Operator : MD/SY
 Sample : 05441-15DL 10X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QK5DL

Quant Time: Nov 22 00:59:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 22 00:29:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	294310	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	277495	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	118103	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	75517	3.430	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.600%
7) Chloroethane-d5	1.907	69	67116	3.976	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.600%
11) 1,1-Dichloroethene-d2	2.563	65	37308	3.840	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.800%
20) 2-Butanone-d5	4.618	46	195705	44.585	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.160%
24) Chloroform-d	5.058	84	178361	4.189	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	5.695	65	93719	4.763	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	5.721	84	364828	4.413	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.682	67	110476	4.434	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.600%
41) Toluene-d8	7.894	98	311217	4.263	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	8.174	79	43303	4.531	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.600%
46) 2-Hexanone-d5	8.627	63	165996	44.200	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.400%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	79579	4.469	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	110887	5.082	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.576	96	6079	0.337	ug/L	# 1
16) Methylene chloride	3.042	84	10213	0.428	ug/L	89
19) 1,1-Dichloroethane	3.859	63	10444	0.284	ug/L	99
29) 1,1,1-Trichloroethane	5.309	97	320319	10.056	ug/L	99
34) Trichloroethene	6.541	95	14169	0.657	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112123\
Data File : VU056565.D
Acq On : 21 Nov 2023 21:36
Operator : MD/SY
Sample : 05441-15DL 10X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QK5DL

Quant Time: Nov 22 00:59:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 22 00:29:26 2023
Response via : Initial Calibration

