

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112123\
 Data File : VU056570.D
 Acq On : 21 Nov 2023 23:37
 Operator : MD/SY
 Sample : 05436-18
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EY1D7

Quant Time: Nov 22 01:00:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 22 00:29:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	260634	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	264164	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	110831	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	78737	4.039	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.907	69	67377	4.508	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.200%
11) 1,1-Dichloroethene-d2	2.563	65	37548	4.364	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.200%
20) 2-Butanone-d5	4.618	46	193160	49.691	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.380%
24) Chloroform-d	5.055	84	182492	4.840	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.695	65	90042	5.168	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	5.721	84	352808	4.483	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	6.682	67	115552	4.871	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.400%
41) Toluene-d8	7.894	98	324970	4.676	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	8.177	79	44218	4.860	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.200%
46) 2-Hexanone-d5	8.627	63	164972	46.145	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.280%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	78503	4.632	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	106047	5.179	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112123\
Data File : VU056570.D
Acq On : 21 Nov 2023 23:37
Operator : MD/SY
Sample : 05436-18
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EY1D7

Quant Time: Nov 22 01:00:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 22 00:29:26 2023
Response via : Initial Calibration

