

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112124\
 Data File : VU061923.D
 Acq On : 21 Nov 2024 17:26
 Operator : MD/SY
 Sample : P4903-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG3Z9

Quant Time: Nov 21 23:10:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 23:06:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	149269	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	149696	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	70266	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	44823	6.122	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.400%
7) Chloroethane-d5	1.907	69	34560	5.336	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.800%
11) 1,1-Dichloroethene-d2	2.560	65	20227	6.915	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	138.200%#
20) 2-Butanone-d5	4.612	46	94471	41.088	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.180%
24) Chloroform-d	5.052	84	88349	4.946	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
26) 1,2-Dichloroethane-d4	5.692	65	45005	4.896	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
32) Benzene-d6	5.718	84	177527	6.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.800%
36) 1,2-Dichloropropane-d6	6.682	67	58013	5.854	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.000%
41) Toluene-d8	7.891	98	154942	6.369	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	127.400%
43) trans-1,3-Dichloroprop...	8.174	79	19005	5.180	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.600%
46) 2-Hexanone-d5	8.624	63	70132	41.880	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.760%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	52285	5.280	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	56704	5.458	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.200%
Target Compounds						
16) Methylene chloride	3.039	84	5704	0.475	ug/L	96
22) cis-1,2-Dichloroethene	4.660	96	6186	0.519	ug/L	98
29) 1,1,1-Trichloroethane	5.303	97	6171	0.358	ug/L	95
34) Trichloroethene	6.534	95	114836	9.518	ug/L	99
42) Toluene	7.968	91	5946	0.125	ug/L	99
47) Tetrachloroethene	8.544	164	483493	55.097	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112124\
Data File : VU061923.D
Acq On : 21 Nov 2024 17:26
Operator : MD/SY
Sample : P4903-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG3Z9

Quant Time: Nov 21 23:10:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 23:06:37 2024
Response via : Initial Calibration

