

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112124\
 Data File : VU061932.D
 Acq On : 21 Nov 2024 21:08
 Operator : MD/SY
 Sample : VIBLK179
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK165

Quant Time: Nov 21 23:13:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 23:06:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	139548	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	131867	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	56139	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	44542	6.507	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	130.200%#
7) Chloroethane-d5	1.907	69	32823	5.421	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.400%
11) 1,1-Dichloroethene-d2	2.560	65	18826	6.884	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	137.600%#
20) 2-Butanone-d5	4.611	46	95985	44.654	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.300%
24) Chloroform-d	5.052	84	83595	5.006	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.692	65	43529	5.065	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.717	84	163804	6.376	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	127.600%#
36) 1,2-Dichloropropane-d6	6.682	67	54399	6.231	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	124.600%
41) Toluene-d8	7.891	98	139627	6.516	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	130.400%#
43) trans-1,3-Dichloroprop...	8.177	79	15400	4.765	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.200%
46) 2-Hexanone-d5	8.624	63	62270	42.212	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.420%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	40556	4.649	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	45677	5.503	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.000%
Target Compounds					Qvalue	
3) Chloromethane	1.518	50	8013	0.637	ug/L	99
16) Methylene chloride	3.036	84	4211	0.375	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112124\
Data File : VU061932.D
Acq On : 21 Nov 2024 21:08
Operator : MD/SY
Sample : VIBLK179
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK165

Quant Time: Nov 21 23:13:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR112024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 23:06:37 2024
Response via : Initial Calibration

