

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045911.D
 Acq On : 22 Nov 2021 14:09
 Operator : SY/MD
 Sample : M4759-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SY-3-20211117

Quant Time: Nov 23 05:14:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	115479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	209683	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	224141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	111625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	90500	52.329	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.660%	
35) Dibromofluoromethane	5.292	113	71775	51.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.540%	
50) Toluene-d8	7.899	98	275558	52.934	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.860%	
62) 4-Bromofluorobenzene	10.632	95	108016	51.745	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.500%	
Target Compounds						
					Qvalue	
16) Acetone	2.658	43	5541	6.294	ug/l	100
52) Toluene	7.970	92	3128	0.844	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045911.D
Acq On : 22 Nov 2021 14:09
Operator : SY/MD
Sample : M4759-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SY-3-20211117

Quant Time: Nov 23 05:14:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

