

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045912.D
 Acq On : 22 Nov 2021 14:31
 Operator : SY/MD
 Sample : M4737-08 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20915

Quant Time: Nov 23 05:14:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	118380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	215430	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	227779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	116280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	92560	52.209	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.420%
35) Dibromofluoromethane	5.292	113	71895	49.986	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	7.899	98	281635	52.658	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.320%
62) 4-Bromofluorobenzene	10.636	95	110262	51.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.820%
Target Compounds						
16) Acetone	2.652	43	6171	6.838	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045912.D
Acq On : 22 Nov 2021 14:31
Operator : SY/MD
Sample : M4737-08 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20915

Quant Time: Nov 23 05:14:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

