

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045919.D
 Acq On : 22 Nov 2021 17:09
 Operator : SY/MD
 Sample : PB140933ZHE#04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140933ZHE#04

Quant Time: Nov 23 05:16:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	113807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	218382	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	231396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	113876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	91190	53.503	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.000%
35) Dibromofluoromethane	5.292	113	72419	49.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.340%
50) Toluene-d8	7.899	98	285566	52.672	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.340%
62) 4-Bromofluorobenzene	10.635	95	109552	50.391	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.780%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.382	74	3506	4.073	ug/l	82
16) Acetone	2.655	43	17115	19.726	ug/l	99
20) Methylene Chloride	3.051	84	6813	4.519	ug/l	99
43) Isopropyl Acetate	5.915	43	13371	3.574	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045919.D
Acq On : 22 Nov 2021 17:09
Operator : SY/MD
Sample : PB140933ZHE#04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140933ZHE#04

Quant Time: Nov 23 05:16:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

