

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045920.D
 Acq On : 22 Nov 2021 17:32
 Operator : SY/MD
 Sample : PB140933ZHE#05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140933ZHE#05

Quant Time: Nov 23 05:17:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	116919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	222462	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	237441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	118014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	92861	53.033	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.060%			
35) Dibromofluoromethane	5.292	113	75064	50.539	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.080%			
50) Toluene-d8	7.899	98	290387	52.578	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.160%			
62) 4-Bromofluorobenzene	10.636	95	112782	50.925	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.840%			
Target Compounds						
8) Diethyl Ether	2.379	74	3665	4.144	ug/l	95
16) Acetone	2.652	43	17158	19.250	ug/l	98
20) Methylene Chloride	3.051	84	6857	4.427	ug/l	93
43) Isopropyl Acetate	5.912	43	13586	3.565	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045920.D
Acq On : 22 Nov 2021 17:32
Operator : SY/MD
Sample : PB140933ZHE#05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140933ZHE#05

Quant Time: Nov 23 05:17:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

