

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045922.D
 Acq On : 22 Nov 2021 18:17
 Operator : SY/MD
 Sample : PB140933ZHE#07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140933ZHE#07

Quant Time: Nov 23 05:17:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	113888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	218038	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	227030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	112743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	90533	53.080	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.160%
35) Dibromofluoromethane	5.292	113	73800	50.696	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.400%
50) Toluene-d8	7.899	98	283768	52.423	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.840%
62) 4-Bromofluorobenzene	10.635	95	108431	49.954	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.900%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.378	74	5542	6.433	ug/l	96
16) Acetone	2.652	43	20275	23.352	ug/l	95
20) Methylene Chloride	3.050	84	5879	3.897	ug/l	96
43) Isopropyl Acetate	5.912	43	14096	3.774	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045922.D
Acq On : 22 Nov 2021 18:17
Operator : SY/MD
Sample : PB140933ZHE#07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140933ZHE#07

Quant Time: Nov 23 05:17:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

