

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045924.D
 Acq On : 22 Nov 2021 19:02
 Operator : SY/MD
 Sample : PB140933ZHE#09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140933ZHE#09

Quant Time: Nov 23 05:18:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	112414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	216036	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	225657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	112203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	90060	53.495	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 106.980%		
35) Dibromofluoromethane	5.292	113	73193	50.745	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 101.500%		
50) Toluene-d8	7.899	98	281809	52.543	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 105.080%		
62) 4-Bromofluorobenzene	10.635	95	108546	50.470	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 100.940%		
Target Compounds						Qvalue
8) Diethyl Ether	2.379	74	3863	4.543	ug/l	95
16) Acetone	2.655	43	16987	19.821	ug/l	96
20) Methylene Chloride	3.051	84	6825	4.583	ug/l	95
43) Isopropyl Acetate	5.912	43	13980	3.777	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045924.D
Acq On : 22 Nov 2021 19:02
Operator : SY/MD
Sample : PB140933ZHE#09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140933ZHE#09

Quant Time: Nov 23 05:18:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

