

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045926.D
 Acq On : 22 Nov 2021 19:47
 Operator : SY/MD
 Sample : M4770-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CC-2

Quant Time: Nov 23 05:19:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	112539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	211720	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	231005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	113916	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	90165	53.498	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.000%
35) Dibromofluoromethane	5.292	113	71934	50.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	7.899	98	282756	53.794	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.580%
62) 4-Bromofluorobenzene	10.635	95	108151	51.312	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.620%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.379	74	5127	6.023	ug/l	94
16) Acetone	2.649	43	16425	19.144	ug/l	94
20) Methylene Chloride	3.051	84	5291	3.549	ug/l	98
43) Isopropyl Acetate	5.915	43	8771	2.418	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045926.D
Acq On : 22 Nov 2021 19:47
Operator : SY/MD
Sample : M4770-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CC-2

Quant Time: Nov 23 05:19:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

