

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045927.D
 Acq On : 22 Nov 2021 20:09
 Operator : SY/MD
 Sample : M4770-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CC-3

Quant Time: Nov 23 05:19:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	115474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	212017	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	231266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	114810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	90554	52.363	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.720%	
35) Dibromofluoromethane	5.292	113	72577	51.272	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.540%	
50) Toluene-d8	7.899	98	287842	54.685	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.380%	
62) 4-Bromofluorobenzene	10.636	95	110864	52.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.060%	
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.382	74	5667	6.488	ug/l	97
16) Acetone	2.655	43	16971	19.278	ug/l	98
20) Methylene Chloride	3.051	84	4965	3.246	ug/l	94
43) Isopropyl Acetate	5.912	43	11074	3.049	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045927.D
Acq On : 22 Nov 2021 20:09
Operator : SY/MD
Sample : M4770-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CC-3

Quant Time: Nov 23 05:19:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

