

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045928.D
 Acq On : 22 Nov 2021 20:32
 Operator : SY/MD
 Sample : M4770-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CC-4

Quant Time: Nov 23 05:19:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	116736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	215512	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	235706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	115922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	92167	52.719	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.440%	
35) Dibromofluoromethane	5.292	113	73442	51.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.080%	
50) Toluene-d8	7.899	98	292177	54.609	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.220%	
62) 4-Bromofluorobenzene	10.636	95	111229	51.843	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.680%	
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.382	74	5373	6.085	ug/l	99
16) Acetone	2.655	43	16309	18.326	ug/l	98
20) Methylene Chloride	3.047	84	5424	3.508	ug/l	86
43) Isopropyl Acetate	5.912	43	8711	2.359	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045928.D
Acq On : 22 Nov 2021 20:32
Operator : SY/MD
Sample : M4770-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CC-4

Quant Time: Nov 23 05:19:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

