

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045935.D
 Acq On : 23 Nov 2021 00:16
 Operator : SY/MD
 Sample : M4775-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PK-10S-20211118

Quant Time: Nov 23 06:50:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	115941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	211042	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	222404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	112141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	87957	50.656	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.320%			
35) Dibromofluoromethane	5.292	113	70576	50.089	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.180%			
50) Toluene-d8	7.899	98	275807	52.641	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.280%			
62) 4-Bromofluorobenzene	10.635	95	105188	50.066	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.140%			
Target Compounds						
16) Acetone	2.658	43	1232	1.394	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045935.D
Acq On : 23 Nov 2021 00:16
Operator : SY/MD
Sample : M4775-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PK-10S-20211118

Quant Time: Nov 23 06:50:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

