

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045940.D
 Acq On : 23 Nov 2021 02:08
 Operator : SY/MD
 Sample : M4775-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PK-10D-20211118

Quant Time: Nov 23 06:51:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	117436	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	213624	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	226364	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	113895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	90037	51.19	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.38%
35) Dibromofluoromethane	5.292	113	71284	49.98	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.96%
50) Toluene-d8	7.899	98	278612	52.53	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.06%
62) 4-Bromofluorobenzene	10.636	95	108628	51.08	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.16%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.385	85	517	0.34	ug/l	93
30) Chloroform	5.092	83	19611	7.35	ug/l	98
44) Trichloroethene	6.546	130	627	0.42	ug/l	82
64) Tetrachloroethene	8.555	164	514	0.40	ug/l	84
65) Chlorobenzene	9.449	112	4346	0.98	ug/l	100
88) 1,4-Dichlorobenzene	11.838	146	1627	0.44	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045940.D
Acq On : 23 Nov 2021 02:08
Operator : SY/MD
Sample : M4775-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PK-10D-20211118

Quant Time: Nov 23 06:51:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

