

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045942.D
 Acq On : 23 Nov 2021 02:52
 Operator : SY/MD
 Sample : M4795-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 HAR-25

Quant Time: Nov 23 06:52:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	116890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	213977	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	228762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	118968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	88442	50.522	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.040%
35) Dibromofluoromethane	5.292	113	70754	49.526	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	7.899	98	282980	53.269	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.540%
62) 4-Bromofluorobenzene	10.632	95	112643	52.879	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.760%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.382	74	5003	5.658	ug/l	100
16) Acetone	2.665	43	27318	30.656	ug/l	98
20) Methylene Chloride	3.050	84	5215	3.368	ug/l	92
43) Isopropyl Acetate	5.912	43	13790	3.762	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045942.D
Acq On : 23 Nov 2021 02:52
Operator : SY/MD
Sample : M4795-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
HAR-25

Quant Time: Nov 23 06:52:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

