

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045944.D
 Acq On : 23 Nov 2021 03:37
 Operator : SY/MD
 Sample : M4795-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 HAR-23

Quant Time: Nov 23 06:52:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	120001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	225185	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	237286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	121765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	90929	50.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.200%
35) Dibromofluoromethane	5.292	113	73264	48.731	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	7.899	98	287406	51.409	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.820%
62) 4-Bromofluorobenzene	10.635	95	114035	50.868	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.740%
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.382	74	5223	5.754	ug/l	98
16) Acetone	2.661	43	23428	25.609	ug/l	99
20) Methylene Chloride	3.051	84	5460	3.435	ug/l	95
43) Isopropyl Acetate	5.912	43	13560	3.515	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045944.D
Acq On : 23 Nov 2021 03:37
Operator : SY/MD
Sample : M4795-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
HAR-23

Quant Time: Nov 23 06:52:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

