

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045929.D
 Acq On : 22 Nov 2021 20:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 05:20:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.648	0.687	-6.0	77	0.00
3 P	Chloromethane	0.742	0.684	7.8	71	0.00
4 C	Vinyl Chloride	0.723	0.735	-1.7#	74	0.00
5 T	Bromomethane	0.403	0.415	-3.0	71	0.01
6 T	Chloroethane	0.472	0.439	7.0	72	0.00
7 T	Trichlorofluoromethane	0.955	0.991	-3.8	77	0.00
8 T	Diethyl Ether	0.378	0.390	-3.2	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.558	0.571	-2.3	76	0.00
10 T	Methyl Iodide	0.730	0.586	19.7	58	0.00
11 T	Tert butyl alcohol					

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
 Data File : VU045929.D
 Acq On : 22 Nov 2021 20:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 05:20:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	66	0.09
50 S	Toluene-d8	1.241	1.202	3.1	73	0.00
51 T	4-Methyl-2-Pentanone	0.565	0.602	-6.5	75	0.00
52 CM	Toluene	0.884	0.936	-5.9#	76	0.00
53 T	t-1,3-Dichloropropene	0.577	0.601	-4.2	74	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.636	-3.8	74	0.00
55 T	1,1,2-Trichloroethane	0.366	0.386	-5.5	77	0.00
56 T	Ethyl methacrylate	0.583	0.599	-2.7	72	0.00
57 T	1,3-Dichloropropane	0.660	0.676	-2.4	75	0.00
58 T	2-Chloroethyl Vinyl ether	0.337	0.308	8.6	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112221\
Data File : VU045929.D
Acq On : 22 Nov 2021 20:54
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Nov 23 05:20:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.069	1.059	0.9	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6