

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112223\
 Data File : VU056598.D
 Acq On : 22 Nov 2023 19:04
 Operator : MD/SY
 Sample : 05441-21
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 23 03:44:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 22 00:29:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	261900	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	255371	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	110464	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	70956	3.622	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.907	69	67337	4.483	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.600%
11) 1,1-Dichloroethene-d2	2.560	65	35540	4.111	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.200%
20) 2-Butanone-d5	4.618	46	208268	53.318	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.640%
24) Chloroform-d	5.055	84	181944	4.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	5.695	65	97472	5.567	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.400%
32) Benzene-d6	5.721	84	353559	4.647	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.685	67	107911	4.706	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.200%
41) Toluene-d8	7.894	98	304745	4.536	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	8.177	79	41213	4.686	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.800%
46) 2-Hexanone-d5	8.627	63	160751	46.512	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.020%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	78777	4.808	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	107100	5.248	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.000%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	9524	0.449	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112223\
Data File : VU056598.D
Acq On : 22 Nov 2023 19:04
Operator : MD/SY
Sample : 05441-21
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 23 03:44:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 22 00:29:26 2023
Response via : Initial Calibration

