




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112321\
Data File : VU045964.D
Acq On : 23 Nov 2021 18:30
Operator : SY/MD
Sample : M4722-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0G12

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 11/24/2021
Supervised By :Semsettin Yesilyurt 11/29/2021

Quant Time: Nov 24 00:33:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 24 00:27:16 2021
Response via : Initial Calibration



TIC: VU045964.D\data.ms

(13) Acetone (T)

2.665min (-0.016) 2.59 ug/L

response 2956

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	35.80	41.58
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112321\
Data File : VU045964.D
Acq On : 23 Nov 2021 18:30
Operator : SY/MD
Sample : M4722-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0G12

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 11/24/2021
Supervised By :Semsettin Yesilyurt 11/29/2021

Quant Time: Nov 24 00:33:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR11521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 24 00:27:16 2021
Response via : Initial Calibration



TIC: VU045964.D\data.ms

(13) Acetone (T)

2.687min (+ 0.006) 7.02 ug/L m

response 8016

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	35.80	15.33
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112321\
 Data File : VU045964.D
 Acq On : 23 Nov 2021 18:30
 Operator : SY/MD
 Sample : M4722-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0G12

Manual IntegrationsAPPROVED

Quant Time: Nov 24 00:33:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 24 00:27:16 2021
 Response via : Initial Calibration

Reviewed By :John Carlone 11/24/2021
 Supervised By :Semsettin Yesilyurt 11/29/2021

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.253	114		91367	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117		92201	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152		41744	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.601	65		17269	2.386	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	47.800%	
7) Chloroethane-d5	1.919	69		26156	4.963	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.200%	
11) 1,1-Dichloroethene-d2	2.572	65		11790	3.883	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.600%	
20) 2-Butanone-d5	4.649	46		131803	67.942	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	135.880%#	
24) Chloroform-d	5.070	84		65362	5.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.400%	
26) 1,2-Dichloroethane-d4	5.706	65		38781	5.480	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.600%	
32) Benzene-d6	5.732	84		132812	5.231	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%	
36) 1,2-Dichloropropane-d6	6.694	67		43059	5.381	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.600%	
41) Toluene-d8	7.903	98		109985	4.788	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%	
43) trans-1,3-Dichloroprop...	8.182	79		12453	3.833	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.600%	
46) 2-Hexanone-d5	8.636	63		91052	62.273	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	124.540%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84		39299	5.813	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152		44354	6.188	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	123.800%#	
Target Compounds							
13) Acetone	2.687	43		8016m	7.020	ug/L	
48) 2-Hexanone	8.694	43		12854	4.006	ug/L #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112321\
Data File : VU045964.D
Acq On : 23 Nov 2021 18:30
Operator : SY/MD
Sample : M4722-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0G12

Manual IntegrationsAPPROVED

Quant Time: Nov 24 00:33:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR11521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 24 00:27:16 2021
Response via : Initial Calibration

Reviewed By :John Carlone 11/24/2021
Supervised By :Semsettin Yesilyurt 11/29/2021

