

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112321\
 Data File : VU045954.D
 Acq On : 23 Nov 2021 14:32
 Operator : SY/MD
 Sample : M4752-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 24 00:31:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 24 00:27:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	96229	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	97433	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	44248	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	36286	4.759	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.200%	
7) Chloroethane-d5	1.916	69	27669	4.985	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.800%	
11) 1,1-Dichloroethene-d2	2.571	65	15089	4.719	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	94.400%	
20) 2-Butanone-d5	4.661	46	104233	51.016	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.040%	
24) Chloroform-d	5.066	84	55315	4.355	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.000%	
26) 1,2-Dichloroethane-d4	5.706	65	35657	4.784	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.600%	
32) Benzene-d6	5.732	84	124362	4.635	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.600%	
36) 1,2-Dichloropropane-d6	6.693	67	39320	4.650	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.000%	
41) Toluene-d8	7.899	98	107252	4.419	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
43) trans-1,3-Dichloroprop...	8.182	79	15726	4.581	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.600%	
46) 2-Hexanone-d5	8.639	63	81416	52.693	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	105.380%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	36088	5.051	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.000%	
66) 1,2-Dichlorobenzene-d4	12.195	152	40469	5.327	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112321\
Data File : VU045954.D
Acq On : 23 Nov 2021 14:32
Operator : SY/MD
Sample : M4752-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 24 00:31:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 24 00:27:16 2021
Response via : Initial Calibration

