

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112322\
 Data File : VU052114.D
 Acq On : 24 Nov 2022 03:24
 Operator : JC/MD
 Sample : N5759-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHB82

Quant Time: Nov 24 07:01:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 24 05:20:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	221762	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	231807	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	88178	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	74743	3.465	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.200%
7) Chloroethane-d5	1.912	69	70730	4.222	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.400%
11) 1,1-Dichloroethene-d2	2.565	65	29741	3.846	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.000%
20) 2-Butanone-d5	4.616	46	211363	54.057	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.120%
24) Chloroform-d	5.060	84	154047	4.860	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	5.700	65	82782	5.171	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	5.726	84	301578	4.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	6.690	67	102026	4.790	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.896	98	228283	3.666	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.400%
43) trans-1,3-Dichloroprop...	8.179	79	31095	4.007	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.629	63	161527	49.110	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.220%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	90154	4.882	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	12.192	152	89907	5.326	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112322\
Data File : VU052114.D
Acq On : 24 Nov 2022 03:24
Operator : JC/MD
Sample : N5759-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHB82

Quant Time: Nov 24 07:01:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 24 05:20:45 2022
Response via : Initial Calibration

