

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112423\
 Data File : VU056602.D
 Acq On : 24 Nov 2023 10:12
 Operator : MD/SY
 Sample : VU1124WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK085

Quant Time: Nov 25 01:38:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 25 01:37:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	290098	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	285640	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	128995	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	93538	4.311	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.907	69	76580	4.603	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.563	65	41113	4.293	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.800%
20) 2-Butanone-d5	4.618	46	198982	45.990	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.980%
24) Chloroform-d	5.055	84	178975	4.264	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	5.698	65	93234	4.807	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	5.724	84	364259	4.280	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.685	67	113997	4.444	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.800%
41) Toluene-d8	7.897	98	299917	3.991	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.180	79	40383	4.105	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.630	63	151583	39.212	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.420%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	75695	4.130	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	103550	4.345	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.000%
Target Compounds						Qvalue
16) Methylene chloride	3.046	84	9656	0.411	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112423\
Data File : VU056602.D
Acq On : 24 Nov 2023 10:12
Operator : MD/SY
Sample : VU1124WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK085

Quant Time: Nov 25 01:38:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 25 01:37:48 2023
Response via : Initial Calibration

