

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112423\
 Data File : VU056603.D
 Acq On : 24 Nov 2023 10:36
 Operator : MD/SY
 Sample : VU1124WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK086

Quant Time: Nov 25 01:38:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 25 01:37:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	262233	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	249569	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	107408	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	94693	4.828	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.600%
7) Chloroethane-d5	1.907	69	71563	4.759	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.563	65	42130	4.867	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.400%
20) 2-Butanone-d5	4.621	46	207614	53.084	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.160%
24) Chloroform-d	5.055	84	188159	4.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.695	65	92490	5.276	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.721	84	357341	4.806	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
36) 1,2-Dichloropropane-d6	6.685	67	117129	5.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.600%
41) Toluene-d8	7.894	98	312279	4.756	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	8.177	79	42882	4.989	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.800%
46) 2-Hexanone-d5	8.631	63	156857	46.441	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.880%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	85985	5.370	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	100939	5.087	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.800%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	8793	0.414	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112423\
Data File : VU056603.D
Acq On : 24 Nov 2023 10:36
Operator : MD/SY
Sample : VU1124WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK086

Quant Time: Nov 25 01:38:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 25 01:37:48 2023
Response via : Initial Calibration

