

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112423\
 Data File : VU056606.D
 Acq On : 24 Nov 2023 12:29
 Operator : MD/SY
 Sample : VIBLK166
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK166

Quant Time: Nov 25 01:39:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 25 01:37:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	241980	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	255750	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	104886	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	93575	5.170	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.400%
7) Chloroethane-d5	1.907	69	68534	4.939	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.800%
11) 1,1-Dichloroethene-d2	2.563	65	42006	5.259	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.200%
20) 2-Butanone-d5	4.618	46	200380	55.522	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.040%
24) Chloroform-d	5.055	84	187442	5.354	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
26) 1,2-Dichloroethane-d4	5.695	65	91991	5.687	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.800%
32) Benzene-d6	5.721	84	355460	4.665	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
36) 1,2-Dichloropropane-d6	6.685	67	116842	5.088	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.800%
41) Toluene-d8	7.894	98	302952	4.502	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
43) trans-1,3-Dichloroprop...	8.177	79	42835	4.863	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.200%
46) 2-Hexanone-d5	8.631	63	172962	49.971	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.940%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	80681	4.917	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	102900	5.311	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.200%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	7360	0.375	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112423\
Data File : VU056606.D
Acq On : 24 Nov 2023 12:29
Operator : MD/SY
Sample : VIBLK166
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK166

Quant Time: Nov 25 01:39:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 25 01:37:48 2023
Response via : Initial Calibration

