

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112618\
Data File : VU028334.D
Acq On : 26 Nov 2018 13:09
Operator : MD/SY
Sample : VU1126WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1126WBL01

Quant Time: Nov 27 01:24:01 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U0112618DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Nov 27 01:03:50 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.74	96	50820	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.31	95	16281	0.87	ug/l	0.00
Spiked Amount 1.000			Recovery	=	87.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	14887	0.86	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.00%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112618\
Data File : VU028334.D
Acq On : 26 Nov 2018 13:09
Operator : MD/SY
Sample : VU1126WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1126WBL01

Quant Time: Nov 27 01:24:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\524U0112618DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Nov 27 01:03:50 2018
Response via : Initial Calibration





#1

Fluorobenzene

Concen: 1.000 ug/l

RT: 5.74 min Scan# 1421

Delta R.T. -0.00 min

Lab File: VU028334.D

Acq: 26 Nov 2018 13:09

Instrument :

MSVOA_U

ClientSampleId :

VU1126WBL01

Tgt Ion: 96 Resp: 50820


Ion Ratio Lower Upper

96 100

70 21.1 17.8 26.8

95 9.3 7.6 11.4

97 6.9 0.0 0.0#






Abundance Ion 96.00 (95.70 to 96.70): VU0

Ion 70.00 (69.70 to 70.70): VU0

Ion 95.00 (94.70 to 95.70): VU0

Ion 97.00 (96.70 to 97.70): VU0



#57

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 10.31 min Scan# 2841

Delta R.T. -0.00 min

Lab File: VU028334.D

Acq: 26 Nov 2018 13:09


Tgt Ion: 95 Resp: 16281

Ion Ratio Lower Upper

95 100

174 73.9 59.3 88.9



176 66.7 54.3 81.5




Abundance Ion 95.00 (94.70 to 95.70): VU0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V






#68
 1,2-Dichlorobenzene-d4
 Concen: Below ug/l
 RT: 11.86 min Scan# 3324
 Delta R.T. 0.00 min
 Lab File: VU028334.D
 Acq: 26 Nov 2018 13:09

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1126WBL01

Tgt Ion:152 Resp: 14887
 Ion Ratio Lower Upper
 152 100
 115 72.5 0.0 146.8
 150 157.8 0.0 653.2



Abundance Ion 152.00 (151.70 to 152.70): V
 Ion 115.00 (114.70 to 115.70): V
 Ion 150.00 (149.70 to 150.70): V

