

Data File : VU035847.D
Acq On : 26 Nov 2019 10:33
Operator : JC/SP
Sample : VU1126WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK14

Quant Time: Nov 27 01:26:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 27 01:18:55 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	237387	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	275557	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	103291	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	171033	5.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.40%
7) Chloroethane-d5	1.93	69	156344	5.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.40%
11) 1,1-Dichloroethene-d2	2.59	63	207034	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	4.70	46	413537	57.62	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.24%
24) Chloroform-d	5.11	84	265630	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.75	65	150831	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.77	84	512449	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
36) 1,2-Dichloropropane-d6	6.73	67	176233	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.60%
41) Toluene-d8	7.93	98	407331	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	8.21	79	61398	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	8.68	63	301499	54.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.84%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	141966	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	12.22	152	122245	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds					Qvalue
13) Acetone	2.69	43	10052	1.595 ug/L	96
34) Trichloroethene	6.58	95	9315	0.330 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU035847.D
Acq On : 26 Nov 2019 10:33
Operator : JC/SP
Sample : VU1126WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK14

Quant Time: Nov 27 01:26:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 27 01:18:55 2019
Response via : Initial Calibration

