

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112619\
 Data File : VU035849.D
 Acq On : 26 Nov 2019 11:37
 Operator : JC/SP
 Sample : K5992-01DL 4X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GARE4DL

Quant Time: Nov 27 01:50:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 27 01:18:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	218608	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	256879	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	71803	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	158249	5.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.80%
7) Chloroethane-d5	1.93	69	143377	5.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.00%
11) 1,1-Dichloroethene-d2	2.60	63	195884	3.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.80%
20) 2-Butanone-d5	4.69	46	250933	37.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.94%
24) Chloroform-d	5.11	84	215540	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.20%
26) 1,2-Dichloroethane-d4	5.75	65	137406	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
32) Benzene-d6	5.77	84	477226	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
36) 1,2-Dichloropropane-d6	6.73	67	162913	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	7.93	98	382235	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.60%
43) trans-1,3-Dichloropropene-	8.21	79	57903	4.50	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.00%
46) 2-Hexanone-d5	8.69	63	245512	47.53	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.06%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	126201	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
64) 1,2-Dichlorobenzene-d4	12.22	152	87517	5.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.20%

Target Compounds					Qvalue
22) cis-1,2-Dichloroethene	4.72	96	13368	0.536 ug/L	95
34) Trichloroethene	6.58	95	153538	5.840 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112619\
Data File : VU035849.D
Acq On : 26 Nov 2019 11:37
Operator : JC/SP
Sample : K5992-01DL 4X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GARE4DL

Quant Time: Nov 27 01:50:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 27 01:18:55 2019
Response via : Initial Calibration

