

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112619\
 Data File : VU035869.D
 Acq On : 26 Nov 2019 20:07
 Operator : JC/SP
 Sample : K6057-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AA1

Quant Time: Nov 27 04:04:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 27 03:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.30	114	216337	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	250323	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	70662	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	144553	5.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.40%
7) Chloroethane-d5	1.94	69	134027	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.60%
11) 1,1-Dichloroethene-d2	2.59	63	176523	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	4.69	46	298673	45.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.34%
24) Chloroform-d	5.12	84	227025	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.75	65	133686	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	5.77	84	450407	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.74	67	152147	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.93	98	343182	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloropropene-	8.22	79	48401	3.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.20%
46) 2-Hexanone-d5	8.69	63	238035	47.29	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.58%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	123132	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
64) 1,2-Dichlorobenzene-d4	12.22	152	82734	5.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.80%

Target Compounds					Qvalue
13) Acetone	2.67	43	19006	3.309 ug/L	91
14) Carbon disulfide	2.83	76	13709	0.183 ug/L	98
34) Trichloroethene	6.58	95	3028	0.118 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112619\
Data File : VU035869.D
Acq On : 26 Nov 2019 20:07
Operator : JC/SP
Sample : K6057-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AA1

Quant Time: Nov 27 04:04:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 27 03:16:13 2019
Response via : Initial Calibration

