

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112619\
 Data File : VU035876.D
 Acq On : 26 Nov 2019 22:58
 Operator : JC/SP
 Sample : K6057-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AA8

Quant Time: Nov 27 04:30:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 27 03:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	206466	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	237626	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	67257	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	139800	5.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.60%
7) Chloroethane-d5	1.93	69	132106	5.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.20%
11) 1,1-Dichloroethene-d2	2.60	63	169403	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	4.69	46	301298	48.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.54%
24) Chloroform-d	5.11	84	222043	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	5.75	65	133136	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
32) Benzene-d6	5.77	84	434880	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.74	67	153674	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.93	98	341533	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.60%
43) trans-1,3-Dichloropropene-	8.21	79	49066	4.12	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.40%
46) 2-Hexanone-d5	8.69	63	263253	55.10	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.20%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	129515	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	12.22	152	81637	5.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.80%

Target Compounds

					Qvalue
13) Acetone	2.66	43	15398	2.809 ug/L	98
14) Carbon disulfide	2.83	76	46324	0.649 ug/L	99
34) Trichloroethene	6.58	95	3527	0.145 ug/L	89
47) Tetrachloroethene	8.58	164	114561	6.190 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112619\
Data File : VU035876.D
Acq On : 26 Nov 2019 22:58
Operator : JC/SP
Sample : K6057-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AA8

Quant Time: Nov 27 04:30:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 27 03:16:13 2019
Response via : Initial Calibration

