

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112718\
 Data File : VU028346.D
 Acq On : 27 Nov 2018 19:25
 Operator : MD/SY
 Sample : J6028-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COMP-2

Quant Time: Nov 28 15:31:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	227694	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	428013	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	386325	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	151561	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	203376	60.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.16%	
35) Dibromofluoromethane	4.88	113	148372	54.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.96%	
50) Toluene-d8	7.57	98	563718	54.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.14%	
62) 4-Bromofluorobenzene	10.30	95	187590	48.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112718\
Data File : VU028346.D
Acq On : 27 Nov 2018 19:25
Operator : MD/SY
Sample : J6028-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COMP-2

Quant Time: Nov 28 15:31:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

