

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU112818\
 Data File : VU028363.D
 Acq On : 28 Nov 2018 14:07
 Operator : MD/SY
 Sample : J6082-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S49-EB-2018-1

Quant Time: Nov 29 06:35:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	237833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	450652	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	405278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	156000	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	210350	52.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.84%	
35) Dibromofluoromethane	4.88	113	150338	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
50) Toluene-d8	7.56	98	585276	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	10.30	95	192605	44.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.18%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.83	59	27407	30.351	ug/l	97
16) Acetone	2.32	43	27371	11.257	ug/l	99
20) Methylene Chloride	2.70	84	44602	12.950	ug/l	97
25) 2-Butanone	4.28	43	5283	1.516	ug/l	99
29) Tetrahydrofuran	4.64	42	189589	83.797	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU112818\
Data File : VU028363.D
Acq On : 28 Nov 2018 14:07
Operator : MD/SY
Sample : J6082-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S49-EB-2018-1

Quant Time: Nov 29 06:35:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

