

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112823\
 Data File : VU056629.D
 Acq On : 28 Nov 2023 15:07
 Operator : MD/SY
 Sample : 05554-03DL 4X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B28DL

Quant Time: Nov 28 23:50:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 28 23:48:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	303557	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	295604	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	128646	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	79316	3.493	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.800%
7) Chloroethane-d5	1.907	69	64690	3.716	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.400%
11) 1,1-Dichloroethene-d2	2.563	65	40441	4.036	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	4.618	46	235152	51.940	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.880%
24) Chloroform-d	5.055	84	202914	4.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
26) 1,2-Dichloroethane-d4	5.695	65	104754	5.162	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	5.721	84	390152	4.430	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.682	67	123566	4.655	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.894	98	338007	4.346	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.000%
43) trans-1,3-Dichloroprop...	8.174	79	45736	4.493	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.800%
46) 2-Hexanone-d5	8.631	63	184846	46.204	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.400%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	92160	4.859	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	119490	5.028	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%
Target Compounds						Qvalue
16) Methylene chloride	3.036	84	9022	0.367	ug/L	89
47) Tetrachloroethene	8.550	164	90969	5.870	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112823\
Data File : VU056629.D
Acq On : 28 Nov 2023 15:07
Operator : MD/SY
Sample : 05554-03DL 4X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B28DL

Quant Time: Nov 28 23:50:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 28 23:48:06 2023
Response via : Initial Calibration

