

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028383.D
 Acq On : 29 Nov 2018 11:08
 Operator : MD/SY
 Sample : J6081-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-EB-2018-3

Quant Time: Nov 30 07:18:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	252310	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	481139	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	438238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	193153	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	219200	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
35) Dibromofluoromethane	4.88	113	160300	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
50) Toluene-d8	7.57	98	628091	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	10.31	95	220831	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
Target Compounds						
16) Acetone	2.32	43	21068	8.168	ug/l	98
20) Methylene Chloride	2.70	84	51695	14.156	ug/l	95
29) Tetrahydrofuran	4.64	42	213556	88.974	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028383.D
Acq On : 29 Nov 2018 11:08
Operator : MD/SY
Sample : J6081-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-EB-2018-3

Quant Time: Nov 30 07:18:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

