

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028389.D
 Acq On : 29 Nov 2018 13:28
 Operator : MD/SY
 Sample : J6081-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0642

Quant Time: Nov 30 11:33:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	247812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	481012	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	450066	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	198852	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	220808	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
35) Dibromofluoromethane	4.88	113	161651	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	7.57	98	636438	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	10.30	95	228089	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
Target Compounds						
16) Acetone	2.32	43	10057	3.970	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028389.D
Acq On : 29 Nov 2018 13:28
Operator : MD/SY
Sample : J6081-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0642

Quant Time: Nov 30 11:33:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

