

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028391.D
 Acq On : 29 Nov 2018 14:14
 Operator : MD/SY
 Sample : J6081-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0640

Quant Time: Nov 30 07:34:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	249134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	466968	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	436039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	202264	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	214936	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	4.88	113	157396	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	7.57	98	615691	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	10.31	95	221264	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
4) Vinyl Chloride	1.40	62	249429	55.032	ug/l	99
11) Tert butyl alcohol	2.83	59	5601	5.921	ug/l #	88
16) Acetone	2.32	43	18997	7.459	ug/l	94
21) trans-1,2-Dichloroethene	2.98	96	5969	1.934	ug/l	96
27) cis-1,2-Dichloroethene	4.23	96	92678	25.388	ug/l	87
95) Naphthalene	13.75	128	8220	3.291	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028391.D
Acq On : 29 Nov 2018 14:14
Operator : MD/SY
Sample : J6081-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0640

Quant Time: Nov 30 07:34:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

