

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028392.D
 Acq On : 29 Nov 2018 14:37
 Operator : MD/SY
 Sample : J6081-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0639

Quant Time: Nov 30 07:35:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	261716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	493796	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	472504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	223494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	230189	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
35) Dibromofluoromethane	4.88	113	166389	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	7.57	98	660439	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	10.31	95	243817	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
Target Compounds						
4) Vinyl Chloride	1.40	62	4555312	956.730	ug/l	99
11) Tert butyl alcohol	2.83	59	2219	2.233	ug/l #	87
12) 1,1-Dichloroethene	2.29	96	2412	0.863	ug/l	88
16) Acetone	2.32	43	15802	5.906	ug/l	96
21) trans-1,2-Dichloroethene	2.99	96	36185	11.159	ug/l	97
27) cis-1,2-Dichloroethene	4.23	96	729384	190.202	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028392.D
Acq On : 29 Nov 2018 14:37
Operator : MD/SY
Sample : J6081-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0639

Quant Time: Nov 30 07:35:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

