

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028394.D
 Acq On : 29 Nov 2018 15:24
 Operator : MD/SY
 Sample : J6081-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0641

Manual Integrations
 APPROVED

MMDadoda
 11/30/2018 11:25:51 AM

Quant Time: Nov 30 07:39:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	256561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	485748	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	457624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	215213	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	225949	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
35) Dibromofluoromethane	4.88	113	163314	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	7.57	98	651535	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	10.30	95	236998	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
Target Compounds						
4) Vinyl Chloride	1.40	62	4790	1.026	ug/l	95
16) Acetone	2.32	43	20872	7.957	ug/l	100
27) cis-1,2-Dichloroethene	4.23	96	2198m	0.585	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028394.D
Acq On : 29 Nov 2018 15:24
Operator : MD/SY
Sample : J6081-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0641

Manual Integrations
APPROVED

MMDadoda
11/30/2018 11:25:51 AM

Quant Time: Nov 30 07:39:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

