

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028397.D
 Acq On : 29 Nov 2018 16:33
 Operator : MD/SY
 Sample : J6085-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 NS-OILY-STONE

Quant Time: Nov 30 07:45:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	249080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	467946	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	444699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	211906	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	220035	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
35) Dibromofluoromethane	4.88	113	156917	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	7.57	98	621907	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	10.30	95	229257	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
Target Compounds						
16) Acetone	2.32	43	22768	8.941	ug/l	98
20) Methylene Chloride	2.70	84	5752076	1603.805	ug/l	100
28) Bromochloromethane	4.56	49	6392	1.894	ug/l #	96
43) Isopropyl Acetate	5.55	43	28746	2.600	ug/l	99
68) m/p-Xylenes	9.38	106	9956	1.590	ug/l	98
95) Naphthalene	13.74	128	104073	8.340	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028397.D
Acq On : 29 Nov 2018 16:33
Operator : MD/SY
Sample : J6085-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
NS-OILY-STONE

Quant Time: Nov 30 07:45:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

