

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
 Data File : VU028400.D
 Acq On : 29 Nov 2018 17:43
 Operator : MD/SY
 Sample : J6134-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-7

Quant Time: Nov 30 07:55:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	255553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	482373	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	468905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	223543	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	223655	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
35) Dibromofluoromethane	4.88	113	165284	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
50) Toluene-d8	7.57	98	657340	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	10.30	95	246394	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
Target Compounds						
16) Acetone	2.32	43	4150	1.588	ug/l #	83
20) Methylene Chloride	2.70	84	2398	0.577	ug/l	96
64) Tetrachloroethene	8.23	164	4364	1.415	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU112918\
Data File : VU028400.D
Acq On : 29 Nov 2018 17:43
Operator : MD/SY
Sample : J6134-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-7

Quant Time: Nov 30 07:55:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

