

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U112919\
 Data File : VU035910.D
 Acq On : 29 Nov 2019 09:58
 Operator : JC/SP
 Sample : VU1129WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK17

Quant Time: Nov 30 03:29:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 30 03:20:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	226585	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	268937	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	102248	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	138285	4.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.40%
7) Chloroethane-d5	1.93	69	134918	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.59	63	172754	3.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.80%
20) 2-Butanone-d5	4.70	46	377250	55.07	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.14%
24) Chloroform-d	5.11	84	251898	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.75	65	139500	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.77	84	469658	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
36) 1,2-Dichloropropane-d6	6.73	67	165137	4.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.60%
41) Toluene-d8	7.93	98	374758	3.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.20%
43) trans-1,3-Dichloropropene-	8.21	79	56825	4.22	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.40%
46) 2-Hexanone-d5	8.68	63	281570	52.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.14%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	141848	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	12.22	152	118861	5.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.00%

Target Compounds				Ovalue
13) Acetone	2.68	43	15905	2.644 ug/L 99
16) Methylene chloride	3.08	84	7252	0.255 ug/L 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU112919\
Data File : VU035910.D
Acq On : 29 Nov 2019 09:58
Operator : JC/SP
Sample : VU1129WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK17

Quant Time: Nov 30 03:29:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 30 03:20:00 2019
Response via : Initial Calibration

